

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004198

Location: GFPD (GRAND FORKS P.D.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. TCS#2 Rep AC 0.100AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 024 Exp. Date 08/01/2014
Test 1 0.082 Test 4 0.081 Test 7 0.079
Test 2 0.081 Test 5 0.081 Test 8 0.081
Test 3 0.081 Test 6 0.081 Test 9 0.081
Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 822 psi Reg 850 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

03 JUN 14
Date
04 June 14 ✓
Date

Charles E. Edr
Field Inspector's Signature
Deb Kasher
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/03/2014 08:56

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 24
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 01/09/2012
Slope 637
Intercept -467208

IR Calibration Date: 01/09/2012
 3um 9um

0th Coef(*100):	-31777	-32014
1st Coef(*100):	271021	132277
2nd Coef(*100):	1454	964
H2O adj(mg/l*10k):	558	576

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT
TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/03/2014 09:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:55
02 Std. Sol.	0.017	09:56
03 Room Air	0.000	09:56
04 Std. Sol.	0.017	09:57
05 Room Air	0.000	09:58
06 Std. Sol.	0.017	09:58
07 Room Air	0.000	09:59


08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 18

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/03/2014 09:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:39
02 Std. Sol.	0.300	09:40
03 Room Air	0.000	09:40
04 Std. Sol.	0.301	09:41
05 Room Air	0.000	09:41
06 Std. Sol.	0.301	09:42
07 Room Air	0.000	09:43

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 13080
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
 0.300 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/03/2014 09:10

Test	AC	Time
01 Room Air	0.000	09:11
02 *Subject Test	INT*	09:12
03 Room Air	0.000	09:12

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 24
Expiration Date = 08/01/2014
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*INTERFERENT
CHECK*

0.100AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/03/2014 08:58

Test	AC	Time
01 Diagnostic	OK	08:59
02 Room Air	RFI*	08:59
03 Room Air	0.000	09:00

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 24
Expiration Date = 08/01/2014
County = 18 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/03/2014 09:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:03
02 Std. Gas	0.082	09:03
03 Room Air	0.000	09:03
04 Std. Gas	0.081	09:04
05 Room Air	0.000	09:04
06 Std. Gas	0.081	09:05
07 Room Air	0.000	09:05

Lot No = 16512080A2
Cyl No = 24
Exp Date = 08/01/2014
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION
CHECK
0.080 AC
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/03/2014 09:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:07
02 Std. Gas	0.081	09:07
03 Room Air	0.000	09:07
04 Std. Gas	0.081	09:08
05 Room Air	0.000	09:08
06 Std. Gas	0.081	09:09
07 Room Air	0.000	09:09

Lot No = 16512080A2

Cyl No = 24

Exp Date = 08/01/2014

County = 18

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION
CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/03/2014 10:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:01
02 Std. Gas	0.079	10:02
03 Room Air	0.000	10:02
04 Std. Gas	0.081	10:03
05 Room Air	0.000	10:03
06 Std. Gas	0.081	10:03
07 Room Air	0.000	10:04


Lot No = 16512080A2

Cyl No = 24

Exp Date = 08/01/2014

County = 18

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*CALIBRATION
CHECK
0.080 AC*

Form 106-I8000