

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004942

Location: UNDP (UND. P.D.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. ICS#2 Rep AC 0.100AC+0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080 A2 Cyl No. 023 Exp. Date 08/01/2014
Test 1 0.081 Test 4 0.080 Test 7 0.080
Test 2 0.079 Test 5 0.080 Test 8 0.079
Test 3 0.081 Test 6 0.080 Test 9 0.080
Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 932 psi Reg 950 psi

K. Remarks/Maintenance record: REPLACED BREATH HOSE O-RINGS,

ANNUAL INSPECTION

03 JUN 14
Date

Charles E. Edm
Field Inspector's Signature

04 June 14 ✓
Date

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = UNDP 8164.13.00 06/09
06/03/2014 11:22

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 23
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 08/21/2012
Slope 638
Intercept -487461

IR Calibration Date: 05/30/2013
 3um 9um

0th Coef(*100): -17966 -17324
1st Coef(*100): 263079 132742
2nd Coef(*100): 2170 1338
H2O adj(mg/l*10k): 723 660

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = UNDP 8164.13.00 06/09
06/03/2014 11:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:32
02 Std. Sol.	0.020	11:33
03 Room Air	0.000	11:33
04 Std. Sol.	0.020	11:34
05 Room Air	0.000	11:35
06 Std. Sol.	0.019	11:35
07 Room Air	0.000	11:36

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 18

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = UNDP 8164.13.00 06/09
06/03/2014 11:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:54
02 Std. Sol.	0.291	11:55
03 Room Air	0.000	11:55
04 Std. Sol.	0.292	11:56
05 Room Air	0.000	11:56
06 Std. Sol.	0.292	11:57
07 Room Air	0.000	11:58


08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 18

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = UNDP 8164.13.00 06/09
06/03/2014 12:00

Test	AC	Time
01 Room Air	0.000	12:00
02 *Subject Test	INT*	12:01
03 Room Air	0.000	12:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 23
Expiration Date = 08/01/2014
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*INTERFERENT
CHECK
0.100AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = UNDP 8164.13.00 06/09
06/03/2014 12:02

Test	AC	Time
01 Diagnostic	OK	12:03
02 Room Air	RFI*	12:03
03 Room Air	0.000	12:03

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 23
Expiration Date = 08/01/2014
County = 18 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = UNDP 8164.13.00 06/09
06/03/2014 12:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:04
02 Std. Gas	0.081	12:04
03 Room Air	0.000	12:05
04 Std. Gas	0.079	12:05
05 Room Air	0.000	12:06
06 Std. Gas	0.081	12:06
07 Room Air	0.000	12:07

Lot No = 16512080A2
Cyl No = 23
Exp Date = 08/01/2014
County = 18 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION
CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = UNDP 8164.13.00 06/09
06/03/2014 12:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:08
02 Std. Gas	0.080	12:08
03 Room Air	0.000	12:09
04 Std. Gas	0.080	12:09
05 Room Air	0.000	12:09
06 Std. Gas	0.080	12:10
07 Room Air	0.000	12:10

Lot No = 16512080A2
Cyl No = 23
Exp Date = 08/01/2014
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION
CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = UNDP 8164.13.00 06/09
06/03/2014 12:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:11
02 Std. Gas	0.080	12:12
03 Room Air	0.000	12:12
04 Std. Gas	0.079	12:12
05 Room Air	0.000	12:13
06 Std. Gas	0.080	12:13
07 Room Air	0.000	12:14

Lot No = 16512080A2
Cyl No = 23
Exp Date = 08/01/2014
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION
CHECK
0.080 AC