

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004947

Location: GRAF (Walsh County S.O.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. ICS#2 Rep AC 0.100AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 046 Exp. Date 08/01/2014
Test 1 0.079 Test 4 0.078 Test 7 0.079
Test 2 0.079 Test 5 0.078 Test 8 0.078
Test 3 0.079 Test 6 0.079 Test 9 0.078
Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 465 psi Reg 475 psi

K. Remarks/Maintenance record: REPLACED BREATH HOSE O-RING,

ANNUAL INSPECTION

02 JUN 14
Date

04 June 14 ✓
Date

Charles E. Egan
Field Inspector's Signature

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = GRAF 8164.13.00 06/09
06/02/2014 14:11

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 46
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 01/19/2010
Slope 679
Intercept -681178

IR Calibration Date: 05/14/2013
 3um 9um

0th Coef(*100): -16611 -19798
1st Coef(*100): 258736 132643
2nd Coef(*100): 2707 1288
H2O adj(mg/l*10k): 677 688

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = GRAF 8164.13.00 06/09
06/02/2014 14:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Sol.	0.017	14:22
03 Room Air	0.000	14:23
04 Std. Sol.	0.018	14:24
05 Room Air	0.000	14:24
06 Std. Sol.	0.019	14:25
07 Room Air	0.000	14:25

08 Sim Temp = 34.0°C

Simul Ser No = DR5115
Std Sol No = 12111
County = 50 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = GRAF 8164.13.00 06/09
06/02/2014 14:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:33
02 Std. Sol.	0.296	14:33
03 Room Air	0.000	14:34
04 Std. Sol.	0.297	14:35
05 Room Air	0.000	14:35
06 Std. Sol.	0.298	14:36
07 Room Air	0.000	14:37


08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 50

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = GRAF 8164.13.00 06/09
06/02/2014 14:17

Test	AC	Time
01 Room Air	0.000	14:18
02 *Subject Test	INT*	14:18
03 Room Air	0.000	14:19

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 46
Expiration Date = 08/01/2014
County = 50 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*INTERFERENT
CHECK
0.100ACT + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = GRAF 8164.13.00 06/09
06/02/2014 14:19

Test	AC	Time
01 Diagnostic	OK	14:20
02 Room Air	RFI*	14:20
03 Room Air	0.000	14:21

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 46
Expiration Date = 08/01/2014
County = 50 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI
 CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = GRAF 8164.13.00 06/09
06/02/2014 14:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:13
02 Std. Gas	0.079	14:13
03 Room Air	0.000	14:14
04 Std. Gas	0.079	14:14
05 Room Air	0.000	14:15
06 Std. Gas	0.079	14:15
07 Room Air	0.000	14:15

Lot No = 16512080A2
Cyl No = 46
Exp Date = 08/01/2014
County = 50 Oper No. = 666666

088 CALIBRATION Charles E. Eder

Operator Signature
CHARLES EDER

Remarks: *CALIBRATION
CHECK*

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = GRAF 8164.13.00 06/09
06/02/2014 14:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Gas	0.078	14:40
03 Room Air	0.000	14:41
04 Std. Gas	0.078	14:41
05 Room Air	0.000	14:42
06 Std. Gas	0.079	14:42
07 Room Air	0.000	14:42

Lot No = 16512080A2
Cyl No = 46
Exp Date = 08/01/2014
County = 50 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION
CHECK
0.080 AC
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = GRAF 8164.13.00 06/09
06/02/2014 14:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:43
02 Std. Gas	0.079	14:44
03 Room Air	0.000	14:44
04 Std. Gas	0.078	14:44
05 Room Air	0.000	14:45
06 Std. Gas	0.078	14:45
07 Room Air	0.000	14:46


Lot No = 16512080A2

Cyl No = 46

Exp Date = 08/01/2014

County = 50

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION
CHECK
0.080 AC

Form 106-I8000