

# CERTIFICATE OF CALIBRATION

## INTOXILYZER® 8000

This is to certify that the calibration of INTOXILYZER serial number 80-003067, manufactured by CMI, subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST (NBS) standard material 1828.

Date 6-5-08 Signed Dale Howard  
Technician

**CMI** INC.

Rec'd 7-28-08  
MAP

316 EAST 9TH STREET  
OWENSBORO, KY. 42303

80-003067  
P. 1 of 1 ym