



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 58282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: SHANNON DUISEN		
Location: TIOGA P.D.		
Intoxilyzer® Serial Number: 80 - 004938		
Gas Lot Number: 709352	Gas Cylinder Number: 21	Expiration Date: 12/23/2011
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0.077 %		
Test 2: 0.077 %		
Test 3: 0.078 %		
(ATTACH TEST RECORD)		

[Signature]
 Chemical Test Operator Signature

6/5/10
 Date

Deb Kashur
 Reviewed By

29 June 2010
 Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TIOG 8164.13.00 06/09
06/05/2010 13:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:47
02 Std. Gas	0.077	13:47
03 Room Air	0.000	13:47
04 Std. Gas	0.077	13:48
05 Room Air	0.000	13:48
06 Std. Gas	0.078	13:49
07 Room Air	0.000	13:49

Lot No = 709352
Cyl No = 21
Exp Date = 12/23/2011
County = 53 Oper No. = 020632



Operator Signature
SHANNON P. DUISEN

Remarks: *ACA MONTHLY TEST*

Form 106-I8000

80-004938
p. 50 of 50
JFM