

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004186

Location: Steele - Kidder Co SO

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5132</u> Lot No. <u> </u> Rep AC <u>0.021</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5134</u> Lot No. <u> </u> Rep AC <u>0.298</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3846</u> Lot No. <u> </u> Rep AC <u>0.10 AC + acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>659357</u>	Cyl No. <u>37</u>	Exp. Date <u>09/03/2011</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 701 psi Reg 700 psi

M. Remarks/Maintenance record: Annual inspection

07 June 2011
Date

Deb Kashur
Field Inspector's Signature

6-8-11
Date

[Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
06/07/2011 14:07

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659357
Standard Cyl #? 37
Standard Expiration? 09/03/2011
Oper No? 777777

Flow Cal. Date: 06/03/2009
Slope 703
Intercept -668211

IR Calibration Date: 06/03/2009

	3um	9um
0th Coef(*100):	-12098	-31908
1st Coef(*100):	257966	136312
2nd Coef(*100):	2942	1009
H2O adj(mg/l*10k):	434	609

***** Printer Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print Test

Form 106-I8000

80-004186
P 107/114
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
06/07/2011 14:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:28
02 Std. Sol.	0.019	14:29
03 Room Air	0.000	14:30
04 Std. Sol.	0.018	14:30
05 Room Air	0.000	14:31
06 Std. Sol.	0.019	14:32
07 Room Air	0.000	14:32


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 22

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

low AC 0.021AC

Form 106-I8000

80-004186
P. 108 of 114
KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
06/07/2011 14:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:36
02 Std. Sol.	0.308	14:37
03 Room Air	0.000	14:37
04 Std. Sol.	0.305	14:38
05 Room Air	0.000	14:39
06 Std. Sol.	0.304	14:39
07 Room Air	0.000	14:40

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 22

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.298 AC

Form 106-I8000

*80-004186
P 109 of 114
vjm*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
06/07/2011 14:33

Test	AC	Time
01 Room Air	0.000	14:34
02 *Subject Test	INT*	14:34
03 Room Air	0.000	14:35

*Invalid Test
Interferent Detected

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10 AC + acetone

Form 106-I8000

80-004186
P. 110 of 114
JEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
06/07/2011 14:23

Test	AC	Time
01 Diagnostic	OK	14:24
02 Room Air	RFI*	14:24
03 Room Air	0.000	14:24

*Invalid Test
Inhibited - RFI

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-004186
p. 111 of 114
JEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
06/07/2011 14:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Gas	0.080	14:10
03 Room Air	0.000	14:10
04 Std. Gas	0.079	14:11
05 Room Air	0.000	14:11
06 Std. Gas	0.079	14:11
07 Room Air	0.000	14:12

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004186
P. 112 of 114
JEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
06/07/2011 14:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:14
02 Std. Gas	0.080	14:14
03 Room Air	0.000	14:15
04 Std. Gas	0.080	14:15
05 Room Air	0.000	14:16
06 Std. Gas	0.080	14:16
07 Room Air	0.000	14:16

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004186
P. 1134114
JEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
06/07/2011 14:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:19
02 Std. Gas	0.080	14:19
03 Room Air	0.000	14:20
04 Std. Gas	0.079	14:20
05 Room Air	0.000	14:20
06 Std. Gas	0.080	14:21
07 Room Air	0.000	14:21

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004186
p 114 of 114
AM