

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004945

Location: McIntosh CoSO - Ashley

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5134 Lot No. Rep AC 0.298 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. Rep AC 0.10 AC + acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. <u>709352</u>	Cyl No. <u>32</u>	Exp. Date <u>12/23/2011</u>
Test 1 <u>0.078</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.078</u>	Test 6 <u>0.078</u>	Test 9 <u>0.079</u>
Average <u>0.078</u>		

L. Tank pressure, Level 3 Func D Sub G P: 712 psi Reg 700 psi

M. Remarks/Maintenance record: Annual inspection, gas cylinder regulator calibration.

07 June 2011

Date

6-9-11

Date

Deb Kashner

Field Inspector's Signature

Deb Kashner

Reviewed by

80-004945
P. 53 of 61
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = ASHL 8164.13.00 06/09
06/07/2011 10:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:45
02 Std. Sol.	0.016	10:45
03 Room Air	0.000	10:46
04 Std. Sol.	0.016	10:47
05 Room Air	0.000	10:47
06 Std. Sol.	0.016	10:48
07 Room Air	0.000	10:48

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 26

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

low AC

Form 106-I8000

*80-004945
P. 55 of 61
km*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = ASHL 8164.13.00 06/09
06/07/2011 10:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Sol.	0.297	11:01
03 Room Air	0.000	11:02
04 Std. Sol.	0.296	11:02
05 Room Air	0.000	11:03
06 Std. Sol.	0.297	11:04
07 Room Air	0.000	11:04

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 26

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.298 AC

Form 106-I8000

*80-004945
P. 56 of 61
FM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = ASHL 8164.13.00 06/09
06/07/2011 11:06

Test	AC	Time
01 Room Air	0.000	11:07
02 *Subject Test	INT*	11:07
03 Room Air	0.000	11:08

*Invalid Test
Interferent Detected

Sub Name = TEST, NA NA
Sub DOB = 08/27/1970
Sub Sex = Unknown Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 709352
Cyl No = 32
Expiration Date = 12/23/2011
County = 26 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent test 0.10 AC + acetone

Form 106-I8000

80-004945
P. 57 of 61
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = ASHL 8164.13.00 06/09
06/07/2011 10:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:20
02 Std. Gas	0.078	10:20
03 Room Air	0.000	10:21
04 Std. Gas	0.078	10:21
05 Room Air	0.000	10:22
06 Std. Gas	0.078	10:22
07 Room Air	0.000	10:23

Lot No = 709352
Cyl No = 32
Exp Date = 12/23/2011
County = 26 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Form 106-I8000

80-004945
P. 59 of 61
KRM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = ASHL 8164.13.00 06/09
06/07/2011 10:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:25
02 Std. Gas	0.079	10:25
03 Room Air	0.000	10:26
04 Std. Gas	0.078	10:26
05 Room Air	0.000	10:26
06 Std. Gas	0.078	10:27
07 Room Air	0.000	10:27

Lot No = 709352
Cyl No = 32
Exp Date = 12/23/2011
County = 26 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

*80-004945
P. 60 of 61
FM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = ASHL 8164.13.00 06/09
06/07/2011 10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:30
02 Std. Gas	0.079	10:30
03 Room Air	0.000	10:30
04 Std. Gas	0.079	10:31
05 Room Air	0.000	10:31
06 Std. Gas	0.079	10:32
07 Room Air	0.000	10:32

Lot No = 709352
Cyl No = 32
Exp Date = 12/23/2011
County = 26 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check ~~0.8~~^{DK} 0.080 AC

Form 106-I8000

80-004945
P. 61 of 61
KAM