

INTOXILYZER 8000 INSPECTION

Serial No.: 80-002666
80-2666 OK

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5144</u> Lot No. <u> </u> Rep AC <u>0.021</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5114</u> Lot No. <u> </u> Rep AC <u>0.298</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5186</u> Lot No. <u> </u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 709352 Cyl No. 6 Exp. Date 23 Dec 2011

Test 1 0.080 Test 4 0.081 Test 7 0.081

Test 2 0.080 Test 5 0.080 Test 8 0.080

Test 3 0.080 Test 6 0.081 Test 9 0.081

Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 916 psi Reg 920 psi

M. Remarks/Maintenance record: Aligned internal printer, annual inspection

08 June 2010
Date

Deb Kashner
Field Inspector's Signature

14 Jun 2010
Date

Mary Pearson
Reviewed by

80-002666
P. 62 of 70
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
06/08/2010 07:51

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 6
Standard Expiration? 12/23/2011
Oper No? 777777

Flow Cal. Date: 09/10/2007
Slope 677
Intercept -603196

IR Calibration Date: 09/10/2007

	3um	9um
0th Coef(*100):	-14148	-8312
1st Coef(*100):	274152	135499
2nd Coef(*100):	1387	855
H2O adj(mg/l*10k):	430	237

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Form 106-18000

80-002666
P. 63 of 70
JAM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
06/08/2010 07:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:56
02 Std. Sol.	0.019	07:57
03 Room Air	0.000	07:57
04 Std. Sol.	0.020	07:58
05 Room Air	0.000	07:59
06 Std. Sol.	0.020	07:59
07 Room Air	0.000	08:00

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

low AC 0.021 AC

Form 106-18000

80-002666
P. 64 of 70
JKH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
06/08/2010 10:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:11
02 Std. Sol.	0.289	10:12
03 Room Air	0.000	10:12
04 Std. Sol.	0.290	10:13
05 Room Air	0.000	10:14
06 Std. Sol.	0.290	10:14
07 Room Air	0.000	10:15

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.298 AC

Form 106-I8000

80-002666
P. L. Sep 70
JCH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
06/08/2010 10:19

Test	AC	Time
01 Room Air	0.000	10:20
02 *Subject Test	INT*	10:20
03 Room Air	0.000	10:21

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 6
Expiration Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 AC + acetone

Form 106-18000

80-002666
P. 46 of 70
JRM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
06/08/2010 10:21

Test	AC	Time
01 Diagnostic	OK	10:22
02 Room Air	RFI*	10:22
03 Room Air	0.000	10:22

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 6
Expiration Date = 12/23/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-18000

80-002666
P. 67 of 70
JK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
06/08/2010 10:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:24
02 Std. Gas	0.080	10:24
03 Room Air	0.000	10:24
04 Std. Gas	0.080	10:25
05 Room Air	0.000	10:25
06 Std. Gas	0.080	10:26
07 Room Air	0.000	10:26

Lot No = 709352

Cyl No = 6

Exp Date = 12/23/2011

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

calibration check 0.080 AC

Form 106-I8000

80-002666
P. G. Prof 70
JTH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
06/08/2010 10:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:28
02 Std. Gas	0.081	10:28
03 Room Air	0.000	10:28
04 Std. Gas	0.080	10:29
05 Room Air	0.000	10:29
06 Std. Gas	0.081	10:30
07 Room Air	0.000	10:30

Lot No = 709352

Cyl No = 6

Exp Date = 12/23/2011

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-18000

*80-002666
P-69d.70
KM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
06/08/2010 10:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Gas	0.081	10:32
03 Room Air	0.000	10:33
04 Std. Gas	0.080	10:33
05 Room Air	0.000	10:33
06 Std. Gas	0.081	10:34
07 Room Air	0.000	10:34

Lot No = 709352

Cyl No = 6

Exp Date = 12/23/2011

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-002666
P. 70-170 *AK*