

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004953  
 06/08/2010 13:48:47

Auto Calibration

pg 1 of 2

<<<<<			3um	>>>>>	<<<<<			9um	>>>>>
-----									
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1									
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)				
Sample #1	0.0690	(-0.0100)		0.1470	(0.0020)				
Sample #2	0.0550	(0.0150)		0.1670	(-0.0110)				
Sample #3	0.0310	(0.0310)		0.1740	(-0.0070)				
Sample #4	0.0830	(0.0180)		0.1770	(-0.0070)				
Avg % Abs	0.0563	(0.0213)		0.1727	(-0.0083)				
STD DEV	0.0260	(0.0085)		0.0051	(0.0023)				
REL STD DEV	46.199	(39.867)		2.972	(27.713)				
-----									
Solution = 0.021 g/210L or 0.1000 mg/l, Samples = 4, Discarded = 1									
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)				
Sample #1	0.3920	(0.0000)		0.8010	(0.0110)				
Sample #2	0.4320	(-0.0110)		0.8540	(-0.0140)				
Sample #3	0.4040	(0.0020)		0.8630	(-0.0180)				
Sample #4	0.4290	(0.0000)		0.8580	(-0.0230)				
Avg % Abs	0.4217	(-0.0030)		0.8583	(-0.0183)				
STD DEV	0.0154	(0.0070)		0.0045	(0.0045)				
REL STD DEV	3.646	(233.333)		0.525	(24.596)				
-----									
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1									
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)				
Sample #1	0.7810	(0.0140)		1.5550	(0.0070)				
Sample #2	0.7870	(0.0210)		1.5260	(0.0210)				
Sample #3	0.7810	(0.0300)		1.5480	(0.0070)				
Sample #4	0.8140	(0.0200)		1.5770	(0.0010)				
Avg % Abs	0.7940	(0.0237)		1.5503	(0.0097)				
STD DEV	0.0176	(0.0055)		0.0256	(0.0103)				
REL STD DEV	2.214	(23.271)		1.650	(106.171)				
-----									
Solution = 0.120 g/210L or 0.5714 mg/l, Samples = 4, Discarded = 1									
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)				
Sample #1	2.2700	(-0.0210)		4.2980	(-0.0190)				
Sample #2	2.2650	(-0.0100)		4.2640	(0.0070)				
Sample #3	2.2350	(0.0120)		4.3050	(0.0050)				
Sample #4	2.2720	(-0.0070)		4.3430	(-0.0100)				
Avg % Abs	2.2573	(-0.0017)		4.3040	(0.0007)				
STD DEV	0.0197	(0.0119)		0.0395	(0.0093)				
REL STD DEV	0.871	(715.821)		0.918	(1393.736)				
-----									
Solution = 0.298 g/210L or 1.4190 mg/l, Samples = 4, Discarded = 1									
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)				
Sample #1	5.2600	(0.0060)		9.7860	(0.0050)				
Sample #2	5.2530	(0.0230)		9.8460	(0.0180)				
Sample #3	5.3110	(0.0230)		9.8960	(-0.0070)				
Sample #4	5.2870	(0.0390)		9.8850	(0.0330)				
Avg % Abs	5.2837	(0.0283)		9.8757	(0.0147)				
STD DEV	0.0291	(0.0092)		0.0263	(0.0202)				
REL STD DEV	0.552	(32.603)		0.266	(137.777)				
-----									

*Calibration*  
*Deb Kashner*  
*p. 1 of 2*  
*lot # 709352*  
*Cyl # 6*  
*Exp date 12/23/11*

80-004953  
 P. 34 of 35 RW

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004953  
 PPPP 0>3A  
 Zero Order Coef -103.43 -182.45  
 PP

Act (g/210L)	Fit (g/210L)	Residual (g/210L)	Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.001	-0.0008	0.000	0.001	-0.0009
0.021	0.020	0.0009	0.021	0.020	0.0009
0.040	0.040	0.0001	0.040	0.040	0.0004
0.120	0.120	-0.0003	0.120	0.120	-0.0004
0.298	0.298	0.0001	0.298	0.298	0.0001

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1

Sample	3um	9um
Sample #1	3105.00	3163.00
Sample #2	3176.00	3182.00
Sample #3	3083.00	3123.00
Sample #4	3079.00	3157.00
Avg	3112.6667	3154.0000
STD DEV	54.8847	29.6142
REL STD DEV	1.763	0.939
H2O adjust (mg/l*10k)	697	655

Atmospheric Pressure = 1010

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

*Calibration*  
*Deb Kashner*  
*p. 2 of 2*  
*lot # 709352*  
*Cyl # 6*  
*Exp date 12/23/11*