



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Richard Johnson</i>		
Location <i>Wash</i>	Intoxilyzer® Serial Number <i>80-004956</i>	
Gas Lot Number <i>09216080A3</i>	Gas Cylinder Number <i>27</i>	Gas Expiration Date <i>05/05/2018</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.080* AC
 B. *0.080* AC
 C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Richard Johnson</i>	Date <i>6/8/2016</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Shaver</i>	Date <i>09 June 2016</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = WASH 8164.13.00 06/09
06/08/2016 14:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:17
02 Std. Gas	0.080	14:17
03 Room Air	0.000	14:18
04 Std. Gas	0.080	14:18
05 Room Air	0.000	14:19
06 Std. Gas	0.081	14:19
07 Room Air	0.000	14:20


Lot No = 09216080A3

Cyl No = 27

Exp Date = 05/05/2018

County = 28

Oper No. = 118707



Operator Signature
RICHARD JOHNSON

Remarks:

Form 106-I8000