

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Phil Wolf

Location: Cavalier, ND

Intoxilyzer Serial No.: 80-002669

Lot Number: 659357

Cylinder Number: 9

Expiration Date: 9/3/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.076 %

Test 2: 0.077 %

Test 3: 0.077 %

(ATTACH TEST RECORD)

P. Wolf
Chemical Test Operator Signature

6/9/10
Date

Deb Kashner
Reviewed By

10 June 2010
Date

80-002669
P. 9/3/11 am


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CAVA 8164.13.00 06/09
06/09/2010 13:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:17
02 Std. Gas	0.076	13:17
03 Room Air	0.000	13:18
04 Std. Gas	0.077	13:18
05 Room Air	0.000	13:18
06 Std. Gas	0.077	13:19
07 Room Air	0.000	13:19

Lot No = 659357
Cyl No = 9
Exp Date = 09/03/2011
County = 34 Oper No. = 120547



Operator Signature
PHIL WOLF

Remarks:

Form 106-I8000

80-002669
P. 93/93 *KA*