



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Barry Vannatta</u>		
Location: <u>Mohl</u>		
Intoxilyzer® Serial Number: <u>80-004947</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>27</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Barry Vannatta
Chemical Test Operator Signature

6-9-11
Date

Deb Kashur
Reviewed By

15 June 2011
Date

80-004947
p. 79 of 80
FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = MOHL 8164.13.00 06/09
06/09/2011 12:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:20
02 Std. Gas	0.080	12:20
03 Room Air	0.000	12:21
04 Std. Gas	0.080	12:21
05 Room Air	0.000	12:22
06 Std. Gas	0.080	12:22
07 Room Air	0.000	12:23

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909



Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004947
P. 80 of 80
DM