

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003061

Location: DEVL (LAKE REG. COR. CEN.)

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3455</u> Lot No. <u>12010</u> Rep AC <u>0.020</u>                                       | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3847</u> Lot No. <u>11110</u> Rep AC <u>0.300</u>                              | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR337</u> Lot No. <u>—</u> Rep AC <u>0.100 AC</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>05911080A1</u>	Cyl No. <u>1</u>	Exp. Date <u>2/1/2013</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.078</u>	Test 8 <u>0.078</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 680 psi Reg 700 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

13 JUNE 12  
Date

Chad E. Egan  
Field Inspector's Signature

15 June 2012  
Date

Deb Kashur  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-003061  
Location = DEVL                   8164.13.00 06/09  
06/13/2012                           18:40

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?            No  
Disable On Memfull?       Yes  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?            0.080  
Standard Lot #?            05911080A1  
Standard Cyl #?            1  
Standard Expiration?       02/01/2013  
Oper No?                    666666

Flow Cal. Date:            03/04/2011  
Slope                       700  
Intercept                  -623726

IR Calibration Date:       04/26/2011  
                            3um                    9um

-----  
0th Coef(\*100):       -10139                   -4793  
1st Coef(\*100):       265984                   136608  
2nd Coef(\*100):       2678                      1277  
H2O adj(mg/l\*10k):    524                        433

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

*PRINT TEST*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003061  
Location = DEVL                    8164.13.00 06/09  
06/13/2012                            19:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:07
02 Std. Sol.	0.020	19:08
03 Room Air	0.000	19:09
04 Std. Sol.	0.020	19:09
05 Room Air	0.000	19:10
06 Std. Sol.	0.021	19:11
07 Room Air	0.000	19:11

08 Sim Temp = 34.0°C

Simul Ser No = 000000  
Std Sol No = 0000  
County = 36                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Low AC*  
*0.020 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer              Alcohol Analyzer  
North Dakota Model 8000              SN 80-003061  
Location = DEVL                      8164.13.00 06/09  
06/13/2012                              19:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:13
02 Std. Sol.	0.296	19:14
03 Room Air	0.000	19:14
04 Std. Sol.	0.298	19:15
05 Room Air	0.000	19:16
06 Std. Sol.	0.299	19:16
07 Room Air	0.000	19:17

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 36

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*LINEARITY TEST*  
*0.300 AC*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer          Alcohol Analyzer  
North Dakota Model 8000          SN 80-003061  
Location = DEVL                  8164.13.00 06/09  
06/13/2012                                  19:00

Test	AC	Time
01 Room Air	0.000	19:01
02 *Subject Test	INT*	19:01
03 Room Air	0.000	19:02

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                  Weight = 150  
Test = DUI                  Cit = QWERTYUIOPASDFG  
Dr. Lic. = ND/DIS821456  
Lot No = 05911080A1  
Cyl No = 1  
Expiration Date = 02/01/2013  
County = 36                  Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*INTERFERENT CHECK*

*0.100AC + ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003061  
Location = DEVL                8164.13.00 06/09  
06/13/2012                        19:03

Test	AC	Time
01 Diagnostic	OK	19:04
02 Room Air	RFI*	19:04
03 Room Air	0.000	19:05

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = MNBVCXZLKJH  
Dr. Lic. = ND/DIS821456  
Lot No = 05911080A1  
Cyl No = 1  
Expiration Date = 02/01/2013  
County = 36                        Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003061  
Location = DEVL                  8164.13.00 06/09  
06/13/2012                            18:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:44
02 Std. Gas	0.079	18:44
03 Room Air	0.000	18:45
04 Std. Gas	0.078	18:45
05 Room Air	0.000	18:46
06 Std. Gas	0.079	18:46
07 Room Air	0.000	18:47

Lot No = 05911080A1  
Cyl No = 1  
Exp Date = 02/01/2013  
County = 36                            Oper No. = 666666

  
\_\_\_\_\_  
Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = DEVL      8164.13.00 06/09  
06/13/2012      18:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:47
02 Std. Gas	0.079	18:48
03 Room Air	0.000	18:48
04 Std. Gas	0.078	18:49
05 Room Air	0.000	18:49
06 Std. Gas	0.079	18:50
07 Room Air	0.000	18:50

Lot No = 05911080A1  
Cyl No = 1  
Exp Date = 02/01/2013  
County = 36      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = DEVL      8164.13.00 06/09  
06/13/2012      18:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:51
02 Std. Gas	0.079	18:51
03 Room Air	0.000	18:52
04 Std. Gas	0.078	18:52
05 Room Air	0.000	18:53
06 Std. Gas	0.078	18:53
07 Room Air	0.000	18:54

Lot No = 05911080A1  
Cyl No = 1  
Exp Date = 02/01/2013  
County = 36      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CEE*  
*CALIBRATION CUT CHECK*  
*0.080AC*