

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003071

Location: COOP (GRIGG'S COUNTY)

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>12010</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>11110</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u> </u> Rep AC <u>0.100AC</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 05911080A1 Cyl No. 5 Exp. Date 02/01/2013
Test 1 0.078 Test 4 0.079 Test 7 0.079
Test 2 0.078 Test 5 0.078 Test 8 0.077
Test 3 0.079 Test 6 0.078 Test 9 0.078
Average 0.078

L. Tank pressure, Level 3 Func D Sub G P: 750 psi Reg 766 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

13 JUNE 12
Date

Charles E. Eden
Field Inspector's Signature

15 June 12
Date

Deb Kasker
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = COOP 8164.13.00 06/09
06/13/2012 15:33

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05911080A1
Standard Cyl #? 5
Standard Expiration? 02/01/2013
Oper No? 666666

Flow Cal. Date: 07/09/2010
Slope 806
Intercept -994737

IR Calibration Date: 04/06/2011
 3um 9um

0th Coef(*100): -10362 -4903
1st Coef(*100): 272212 136536
2nd Coef(*100): 2666 1129
H2O adj(mg/l*10k): 537 411

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINTTEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = COOP 8164.13.00 06/09
06/13/2012 15:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:52
02 Std. Sol.	0.020	15:53
03 Room Air	0.000	15:53
04 Std. Sol.	0.020	15:54
05 Room Air	0.000	15:55
06 Std. Sol.	0.020	15:55
07 Room Air	0.000	15:56

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 20

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = COOP 8164.13.00 06/09
06/13/2012 15:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:57
02 Std. Sol.	0.294	15:58
03 Room Air	0.000	15:59
04 Std. Sol.	0.295	15:59
05 Room Air	0.000	16:00
06 Std. Sol.	0.295	16:01
07 Room Air	0.000	16:01

08 Sim Temp = 34.0°C

Simul Ser No = 000000
Std Sol No = 0000
County = 20 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LINEARITY TEST*
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = COOP 8164.13.00 06/09
06/13/2012 16:02

Test	AC	Time
01 Room Air	0.000	16:03
02 *Subject Test	INT*	16:04
03 Room Air	0.000	16:04

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJHG
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 20 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100 AC + ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = COOP 8164.13.00 06/09
06/13/2012 15:35

Test	AC	Time
01 Diagnostic	OK	15:36
02 Room Air	RFI*	15:36
03 Room Air	0.000	15:37

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 05911080A1
Cyl No = 5
Expiration Date = 02/01/2013
County = 20 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = COOP 8164.13.00 06/09
06/13/2012 15:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:38
02 Std. Gas	0.078	15:39
03 Room Air	0.000	15:39
04 Std. Gas	0.078	15:39
05 Room Air	0.000	15:40
06 Std. Gas	0.079	15:40
07 Room Air	0.000	15:41

Lot No = 05911080A1
Cyl No = 5
Exp Date = 02/01/2013
County = 20 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = COOP 8164.13.00 06/09
06/13/2012 15:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:42
02 Std. Gas	0.079	15:42
03 Room Air	0.000	15:43
04 Std. Gas	0.078	15:43
05 Room Air	0.000	15:44
06 Std. Gas	0.078	15:44
07 Room Air	0.000	15:45

Lot No = 05911080A1
Cyl No = 5
Exp Date = 02/01/2013
County = 20 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK

0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = COOP 8164.13.00 06/09
06/13/2012 15:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:45
02 Std. Gas	0.079	15:46
03 Room Air	0.000	15:46
04 Std. Gas	0.077	15:47
05 Room Air	0.000	15:47
06 Std. Gas	0.078	15:48
07 Room Air	0.000	15:48

Lot No = 05911080A1
Cyl No = 5
Exp Date = 02/01/2013
County = 20 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000