



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Danielle Rosevaren</u>		
Location: <u>CARR</u>		
Intoxilyzer® Serial Number: <u>80-004944</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>12</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <u>0.078</u> %		
Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

Danielle Rosevaren  
Chemical Test Operator Signature

06/19/10  
Date

Deb Kashner  
Reviewed By

17 June 2010  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004944  
Location = CARR      8164.13.00 06/09  
06/14/2010      17:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:07
02 Std. Gas	0.079	17:07
03 Room Air	0.000	17:08
04 Std. Gas	0.078	17:08
05 Room Air	0.000	17:08
06 Std. Gas	0.079	17:09
07 Room Air	0.000	17:09

Lot No = 709352  
Cyl No = 12  
Exp Date = 12/23/2011  
County = 16      Oper No. = 100754

*Danielle Roseware*

Operator Signature  
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004944  
P. 50 of 50  
jam