

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004194

Location: LANG (CAVALIER Co. S.O.)

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3455</u> Lot No. <u>12010</u> Rep AC <u>0.020</u>                                     | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3847</u> Lot No. <u>11110</u> Rep AC <u>0.300</u>                            | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3377</u> Lot No. <u>—</u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>23411080A1</u>	Cyl No. <u>4</u>	Exp. Date <u>10/1/2013</u>
Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 984 psi Reg 1000 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

14 JUNE 12  
Date

[Signature]  
Field Inspector's Signature

15 June 12  
Date

Deb Kashner  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004194  
Location = LANG                    8164.13.00 06/09  
06/14/2012                            12:45

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            04  
Standard Expiration?       10/01/2013  
Oper No?                    666666

Flow Cal. Date:            06/23/2009  
Slope                      681  
Intercept                  -622430

IR Calibration Date:       06/29/2011  
                                    3um                      9um

-----  
0th Coef(\*100):            1196                      -615  
1st Coef(\*100):            247548                    123137  
2nd Coef(\*100):            7878                      2577  
H2O adj(mg/l\*10k):       583                        655

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004194  
Location = LANG                    8164.13.00 06/09  
06/14/2012                            12:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Sol.	0.019	12:58
03 Room Air	0.000	12:59
04 Std. Sol.	0.020	12:59
05 Room Air	0.000	13:00
06 Std. Sol.	0.021	13:01
07 Room Air	0.000	13:01

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 10

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Low AC*

*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = LANG      8164.13.00 06/09  
06/14/2012      13:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:04
02 Std. Sol.	0.298	13:05
03 Room Air	0.000	13:05
04 Std. Sol.	0.304	13:06
05 Room Air	0.000	13:06
06 Std. Sol.	0.306	13:07
07 Room Air	0.000	13:08


08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 10

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

LINEARITY TEST  
0.300 AC

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004194  
Location = LANG                8164.13.00 06/09  
06/14/2012                        12:52

Test	AC	Time
01 Room Air	0.000	12:54
02 *Subject Test	INT*	12:54
03 Room Air	0.000	12:55

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = MNBVCXZLKJHG  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 04  
Expiration Date = 10/01/2013  
County = 10                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*INTERFERENT CHECK  
0.100 AC + ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004194  
Location = LANG                8164.13.00 06/09  
06/14/2012                      12:50

Test	AC	Time
01 Diagnostic	OK	12:51
02 Room Air	RFI*	12:51
03 Room Air	0.000	12:52

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = QWERTYUIOPASDFG  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 04  
Expiration Date = 10/01/2013  
County = 10                              Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = LANG      8164.13.00 06/09  
06/14/2012      13:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Gas	0.081	13:11
03 Room Air	0.000	13:12
04 Std. Gas	0.081	13:12
05 Room Air	0.000	13:13
06 Std. Gas	0.080	13:13
07 Room Air	0.000	13:13

Lot No = 23411080A1  
Cyl No = 4  
Exp Date = 10/01/2013  
County = 10      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = LANG      8164.13.00 06/09  
06/14/2012      13:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:14
02 Std. Gas	0.081	13:15
03 Room Air	0.000	13:15
04 Std. Gas	0.081	13:15
05 Room Air	0.000	13:16
06 Std. Gas	0.081	13:16
07 Room Air	0.000	13:17

Lot No = 23411080A1  
Cyl No = 4  
Exp Date = 10/01/2013  
County = 10      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

0.080 AC  
CALIBRATION CHECK

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004194  
Location = LANG                    8164.13.00 06/09  
06/14/2012                            13:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:18
02 Std. Gas	0.081	13:18
03 Room Air	0.000	13:19
04 Std. Gas	0.081	13:19
05 Room Air	0.000	13:20
06 Std. Gas	0.081	13:20
07 Room Air	0.000	13:21

Lot No = 23411080A1  
Cyl No = 4  
Exp Date = 10/01/2013  
County = 10                            Oper No. = 666666

  
Operator Signature  
CHARLES EDER

Remarks: