



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>MARK PEARSON</i>		
Location <i>EMMONS COUNTY</i>	Intoxilyzer® Serial Number <i>80-004935</i>	
Gas Lot Number <i>34921080A1</i>	Gas Cylinder Number <i>56</i>	Gas Expiration Date <i>2/5/24</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.017* AC
 B. *0.015* AC
 C. *0.026* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

5. Send the Following to the Crime Laboratory: *Note: All forms were received and reviewed at the Crime Laboratory on 6/14/22.*
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G). *20 July 22*
 B. ACA Test Record. *30 July 22*

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>6/14/22</i>
Reviewed By (Crime Laboratory Use Only) <i>Janelle Pertschilker</i>	Date <i>6/14/22</i>

Intoxilyzer Test Record and Checklist
 NDCS Crime Lab, Div., Bismarck, ND 58501
 OTC, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN: 83-004935
 Location = L1N 6:54:41.0 09/16
 05/14/2022 12:35

DRY CAL CHECK

Test	AC	Test
01 Room Air	0.00	12:00
02 S10. S85	0.077	12:05
03 Room Air	0.010	12:10
04 S10. S85	0.075	12:15
05 Room Air	0.000	12:20
06 S10. S85	0.076	12:25
07 Room Air	0.000	12:30

Lot No = 34921080A1
 C.I. No = 56
 Exp Date = 02/05/2024
 C001y = 15
 Oper No. = 130117

[Signature]
 Operator Signature
 MARK PEARSON

Remarks:

Form 106-18000