



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>100414</u>		
Location: <u>Stutsman Co.</u>		
Intoxilyzer® Serial Number: <u>80-004198</u>		
Gas Lot Number: <u>05911080A1</u>	Gas Cylinder Number: <u>21</u>	Expiration Date: <u>02/01/2013</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.077</u> % Test 2: <u>0.077</u> % Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

[Signature]
 Chemical Test Operator Signature

Deb Kashner
 Reviewed By

06/15/2011
 Date

29 June 2011
 Date

80-004198
 P. 138 of 139
 km


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = JAME 8164.13.00 06/09
06/15/2011 10:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:09
02 Std. Gas	0.077	10:09
03 Room Air	0.000	10:10
04 Std. Gas	0.077	10:10
05 Room Air	0.000	10:11
06 Std. Gas	0.078	10:11
07 Room Air	0.000	10:11

Lot No = 05911080A1
Cyl No = 21
Exp Date = 02/01/2013
County = 47 Oper No. = 100414


Operator Signature
JOSHUA RUDE

Remarks:

Form 106-I8000

80-004198
P. 139 of 139
van