



# SERVICE WORK ORDER

316 E. 9<sup>th</sup> St. Owensboro KY 42303  
Phone 866-835-0690  
Fax: 270-685-6268

336749

DATE	8-15-15
RCVD VIA	FEUX

<b>BILL TO:</b>		
ATTN:		
CITY	STATE	ZIP
CUSTOMER NO.		
PHONE		

<b>SHIP TO:</b>		
ATTN:		
CITY	STATE	ZIP
PHONE		
FAX		

MODEL #	SERIAL #	BILL CODE
8000	80-005746	EW
EXTRA PARTS RCVD		
1BT gas box, battery		
DESCRIPTION OF PROBLEM		
External Battery issue		

**Infrared / Fuel Cell**  
PRELIMINARY TEST INFORMATION

BATTERY-EXT OK V. INT 3.05V

VERIFIED PROBLEM?

12V BTY would not hold charge - yes

WORK PERFORMED:

replaced 12V BTY

charged new BTY-12V- OK

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	013010	12V BTY

**Miscellaneous**  
PRELIMINARY TEST INFORMATION

8164.13

WORK PERFORMED:

100

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
DU	NDP	NDP	NDP
1/2 HOURS	HOURS	HOURS	HOURS
10	0.5	1.0	1.0
DATE	DATE	DATE	DATE
Aug 5, 2015	Aug 10, 15	Aug 11, 15	Aug 11, 15

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS