



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Dean Haaland				
Location:	Cass				
Intoxilyzer® Serial Number:	80-004205				
Gas Lot Number:	709352	Gas Cylinder Number:	14	Expiration Date:	12-23-11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0. <u>080</u> %					
Test 2: 0. <u>080</u> %					
Test 3: 0. <u>080</u> %					
(ATTACH TEST RECORD)					

*Dean Haaland*  
Chemical Test Operator Signature

6-16-10  
Date

Deb Kashner  
Reviewed By

17 June 2010  
Date

80-004205  
p. 81 of 82  
FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004205  
Location = CASS      8164.13.00 06/09  
06/16/2010      11:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:16
02 Std. Gas	0.080	11:17
03 Room Air	0.000	11:17
04 Std. Gas	0.080	11:18
05 Room Air	0.000	11:18
06 Std. Gas	0.080	11:18
07 Room Air	0.000	11:19

Lot No = 709352  
Cyl No = 14  
Exp Date = 12/23/2011  
County = 09      Oper No. = 109409




Operator Signature  
DEAN J. HAALAND

Remarks:

Form 106-I8000

I hereby certify that the information in the  
attached document is a true and correct copy  
of the original document received by this office



Deputy Sheriff  
Cass County Sheriff's Office

80-004205  
P. 82 of 82  
KAM