



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Tyler A Wolfe</u>		
Location: <u>HARV</u>		
Intoxilyzer® Serial Number: <u>80-004935</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>45</u>	Expiration Date: <u>12/23/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.075</u> %		
Test 2: <u>0.076</u> %		
Test 3: <u>0.075</u> %		
(ATTACH TEST RECORD)		

Tyler Wolfe
Chemical Test Operator Signature

6/16/10
Date

Deb Kashur
Reviewed By

21 June 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = HARV 8164.13.00 06/09
06/16/2010 17:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:58
02 Std. Gas	0.075	17:58
03 Room Air	0.000	17:59
04 Std. Gas	0.076	17:59
05 Room Air	0.000	18:00
06 Std. Gas	0.075	18:00
07 Room Air	0.000	18:00

Lot No = 709352
Cyl No = 45
Exp Date = 12/23/2011
County = 52 Oper No. = 060901



Operator Signature
TYLER A. WOLFE

Remarks:

ACA

Form 106-I8000

80-004935
P. 55 of 55
KLM