



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>William Petty</u>		
Location: <u>GFAE</u>		
Intoxilyzer® Serial Number: <u>80-004949</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>39</u>	Expiration Date: <u>09/03/11</u>
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.076</u> % Test 2: <u>0.076</u> % Test 3: <u>0.076</u> %		
(ATTACH TEST RECORD) <u>ACV</u>		

Will. Petty
Chemical Test Operator Signature

16 Jun 10
Date

Deb Kashur
Reviewed By

30 June 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004949
Location = GFAP 8164.13.00 06/09
06/16/2010 10:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:47
02 Std. Gas	0.076	10:47
03 Room Air	0.000	10:48
04 Std. Gas	0.076	10:48
05 Room Air	0.000	10:49
06 Std. Gas	0.076	10:49
07 Room Air	0.000	10:50

Lot No = 659358
Cyl No = 39
Exp Date = 09/03/2011
County = 18 Oper No. = 120011



Operator Signature
WILLIAM PETTY

Remarks: *ACA*

Form 106-I8000

80-004949
P. 44-44
km