

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003067

Location: Truill Co SO

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3845</u> Lot No. <u>13070</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3841</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5191</u> Lot No. <u>1CS#1</u> Rep AC <u>0.100 AC + acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 23411080A1 Cyl No. 37 Exp. Date 10/01/2013

Test 1 <u>0.079</u>	Test 4 <u>0.081</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>

Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 940 psi Reg 950 psi

M. Remarks/Maintenance record: Annual inspection

11 June 13
Date
17 JUN 13
Date

Deb Kashur
Field Inspector's Signature
Charles E. Edr
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
06/11/2013 15:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:56
02 Std. Sol.	0.017	15:57
03 Room Air	0.000	15:58
04 Std. Sol.	0.017	15:59
05 Room Air	0.000	15:59
06 Std. Sol.	0.017	16:00
07 Room Air	0.000	16:00

08 Sim Temp = 34.0°C

Simul Ser No = DR3845

Std Sol No = 13070

County = 49

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
06/11/2013 15:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:51
02 Std. Sol.	0.297	15:51
03 Room Air	0.000	15:52
04 Std. Sol.	0.297	15:53
05 Room Air	0.000	15:53
06 Std. Sol.	0.296	15:54
07 Room Air	0.000	15:54

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 13020

County = 49

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks: *DK*

e Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
06/11/2013 16:02

Test	AC	Time
01 Room Air	0.000	16:02
02 *Subject Test	INT*	16:03
03 Room Air	0.000	16:03

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 37
Expiration Date = 10/01/2013
County = 49 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
06/11/2013 15:36

Test	AC	Time
01 Diagnostic	OK	15:37
02 Room Air	RFI*	15:37
03 Room Air	0.000	15:37

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 37
Expiration Date = 10/01/2013
County = 49 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
06/11/2013 15:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:39
02 Std. Gas	0.079	15:39
03 Room Air	0.000	15:39
04 Std. Gas	0.080	15:40
05 Room Air	0.000	15:40
06 Std. Gas	0.080	15:41
07 Room Air	0.000	15:41

Lot No = 23411080A1
Cyl No = 37
Exp Date = 10/01/2013
County = 49 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
06/11/2013 15:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:42
02 Std. Gas	0.081	15:42
03 Room Air	0.000	15:43
04 Std. Gas	0.079	15:43
05 Room Air	0.000	15:44
06 Std. Gas	0.079	15:44
07 Room Air	0.000	15:45

Lot No = 23411080A1
Cyl No = 37
Exp Date = 10/01/2013
County = 49 Oper No. = 777777

DK
Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
06/11/2013 15:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:46
02 Std. Gas	0.080	15:46
03 Room Air	0.000	15:47
04 Std. Gas	0.079	15:47
05 Room Air	0.000	15:48
06 Std. Gas	0.080	15:48
07 Room Air	0.000	15:48

Lot No = 23411080A1
Cyl No = 37
Exp Date = 10/01/2013
County = 49 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC