

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004198

Location: GFPD

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3845</u> Lot No. <u>13670</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3841</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5191</u> Lot No. <u>KS#1</u> Rep AC <u>0.100 + acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>62</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.081</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 742 psi Reg 750 psi

M. Remarks/Maintenance record: Annual inspection, replaced o-ring on breath hose.

11 June 13
Date 17
13 JUN 13
Date CEE

Deb Kashur
Field Inspector's Signature
Charles E. Edwards
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/11/2013 11:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:32
02 Std. Sol.	0.016	11:33
03 Room Air	0.000	11:34
04 Std. Sol.	0.016	11:34
05 Room Air	0.000	11:35
06 Std. Sol.	0.016	11:36
07 Room Air	0.000	11:36

08 Sim Temp = 34.0°C

Simul Ser No = DR3845

Std Sol No = 13070

County = 18

Oper No. = 777777

Deb Kasher

Operator Signature

DEB KASHUR

Remarks:

low AC 0.020 AC
~~0.300 AC~~ OK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/11/2013 11:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:38
02 Std. Sol.	0.296	11:39
03 Room Air	0.000	11:40
04 Std. Sol.	0.296	11:40
05 Room Air	0.000	11:41
06 Std. Sol.	0.294	11:42
07 Room Air	0.000	11:42

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 13020

County = 18

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/11/2013 11:43

Test	AC	Time
01 Room Air	0.000	11:44
02 *Subject Test	INT*	11:45
03 Room Air	0.000	11:45

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 62
Expiration Date = 10/01/2013
County = 18 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

*Interferent check 0.100 AC
+ acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/11/2013 11:16

Test	AC	Time
01 Diagnostic	OK	11:16
02 Room Air	RFI*	11:16
03 Room Air	0.000	11:17

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 62
Expiration Date = 10/01/2013
County = 18 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/11/2013 11:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Gas	0.081	11:18
03 Room Air	0.000	11:19
04 Std. Gas	0.080	11:19
05 Room Air	0.000	11:19
06 Std. Gas	0.080	11:20
07 Room Air	0.000	11:20

Lot No = 23411080A1

Cyl No = 62

Exp Date = 10/01/2013

County = 18

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/11/2013 11:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.080	11:22
03 Room Air	0.000	11:22
04 Std. Gas	0.080	11:22
05 Room Air	0.000	11:23
06 Std. Gas	0.080	11:23
07 Room Air	0.000	11:24

Lot No = 23411080A1

Cyl No = 62

Exp Date = 10/01/2013

County = 18

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 Ae

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
06/11/2013 11:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:26
02 Std. Gas	0.081	11:27
03 Room Air	0.000	11:27
04 Std. Gas	0.080	11:27
05 Room Air	0.000	11:28
06 Std. Gas	0.080	11:28
07 Room Air	0.000	11:29

Lot No = 23411080A1
Cyl No = 62
Exp Date = 10/01/2013
County = 18 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080AC