

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004203

Location: Rugby - Heart of America Correction

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DR3845 Lot No. 13670 Rep AC 0.020 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DR3841 Lot No. 13020 Rep AC 0.300 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DR5191 Lot No. 105#1 Rep AC 0.100 AC + acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. 2344080A1 Cyl No. 54 Exp. Date 10/01/2013  
Test 1 0.080 Test 4 0.081 Test 7 0.080  
Test 2 0.079 Test 5 0.080 Test 8 0.080  
Test 3 0.080 Test 6 0.080 Test 9 0.080  
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 982 psi Reg 990 psi

M. Remarks/Maintenance record: Annual inspection, replaced ext. battery & o-rings on breath hose & simulator return.

10 June 13  
Date  
17 JUN 13  
Date

Deb Keshner  
Field Inspector's Signature  
Charles E. Ehr  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000          SN 80-004203  
Location = RUGB                    8164.13.00 06/09  
06/10/2013                            10:56

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            54  
Standard Expiration?       10/01/2013  
Oper No?                    777777

Flow Cal. Date:            01/28/2010  
Slope                      719  
Intercept                  -711903

IR Calibration Date:        01/28/2010

	3um	9um
0th Coef(*100):	-17469	-28710
1st Coef(*100):	264216	135022
2nd Coef(*100):	2778	1081
H2O adj(mg/l*10k):	430	605

\*\*\*\*\* Printer Test End \*\*\*\*\*

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
06/10/2013      11:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:09
02 Std. Sol.	0.016	11:10
03 Room Air	0.000	11:11
04 Std. Sol.	0.016	11:11
05 Room Air	0.000	11:12
06 Std. Sol.	0.016	11:12
07 Room Air	0.000	11:13

08 Sim Temp = 34.0°C

Simul Ser No = DR3845  
Std Sol No = 13070  
County = 35      Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

Low AC 0.020 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
06/10/2013      11:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:49
02 Std. Sol.	0.285	11:50
03 Room Air	0.000	11:51
04 Std. Sol.	0.285	11:51
05 Room Air	0.000	11:52
06 Std. Sol.	0.287	11:53
07 Room Air	0.000	11:53

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 13020

County = 35

Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
06/10/2013      11:23

Test	AC	Time
01 Room Air	0.000	11:24
02 *Subject Test	INT*	11:24
03 Room Air	0.000	11:25

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 54  
Expiration Date = 10/01/2013  
County = 35      Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004203  
Location = RUGB                    8164.13.00 06/09  
06/10/2013                            10:57

Test	AC	Time
01 Diagnostic	OK	10:58
02 Room Air	RFI*	10:58
03 Room Air	0.000	10:58

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 54  
Expiration Date = 10/01/2013  
County = 35                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Keshur*

Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
06/10/2013      10:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:50
02 Std. Gas	0.080	10:51
03 Room Air	0.000	10:51
04 Std. Gas	0.079	10:51
05 Room Air	0.000	10:52
06 Std. Gas	0.080	10:52
07 Room Air	0.000	10:53

Lot No = 23411080A1  
Cyl No = 54  
Exp Date = 10/01/2013  
County = 35      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004203  
Location = RUGB                8164.13.00 06/09  
06/10/2013                        10:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:59
02 Std. Gas	0.081	11:00
03 Room Air	0.000	11:00
04 Std. Gas	0.080	11:01
05 Room Air	0.000	11:01
06 Std. Gas	0.080	11:02
07 Room Air	0.000	11:02

Lot No = 23411080A1  
Cyl No = 54  
Exp Date = 10/01/2013  
County = 35                            Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080 AC*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004203  
Location = RUGB                8164.13.00 06/09  
06/10/2013                        11:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Gas	0.080	11:21
03 Room Air	0.000	11:21
04 Std. Gas	0.080	11:22
05 Room Air	0.000	11:22
06 Std. Gas	0.080	11:22
07 Room Air	0.000	11:23

Lot No = 23411080A1  
Cyl No = 54  
Exp Date = 10/01/2013  
County = 35                            Oper No. = 777777

*Deb Keshur*  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080AC*