

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004954 Location: Rolle He Co SO

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3845</u> Lot No. <u>13070</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3841</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5191</u> Lot No. <u>1CS#1</u> Rep AC <u>0.10 AC + acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 23411080A1 Cyl No. 61 Exp. Date 10/01/2013
Test 1 0.078 Test 4 0.080 Test 7 0.080
Test 2 0.079 Test 5 0.080 Test 8 0.079
Test 3 0.078 Test 6 0.080 Test 9 0.080
Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 813 psi Reg 820 psi

M. Remarks/Maintenance record: annual inspection
10
~~06~~ June 2013
Date
17 JUN 13
Date
Deb Kasher
Field Inspector's Signature
Charles E. E...
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
06/10/2013 14:19

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 61
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 08/27/2012
Slope 670
Intercept -570610

IR Calibration Date: 08/27/2012
 3um 9um

0th Coef(*100):	-22689	-21316
1st Coef(*100):	260406	132331
2nd Coef(*100):	1084	919
H2O adj(mg/l*10k):	570	470

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
06/10/2013 14:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Sol.	0.017	14:38
03 Room Air	0.000	14:39
04 Std. Sol.	0.018	14:39
05 Room Air	0.000	14:40
06 Std. Sol.	0.018	14:41
07 Room Air	0.000	14:41

08 Sim Temp = 34.0°C

Simul Ser No = DR3845

Std Sol No = 13070

County = 40

Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
06/10/2013 14:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:44
02 Std. Sol.	0.285	14:45
03 Room Air	0.000	14:45
04 Std. Sol.	0.287	14:46
05 Room Air	0.000	14:46
06 Std. Sol.	0.288	14:47
07 Room Air	0.000	14:48

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 13020

County = 40

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
06/10/2013 14:53

Test	AC	Time
01 Room Air	0.000	14:54
02 *Subject Test	INT*	14:54
03 Room Air	0.000	14:55

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 61
Expiration Date = 10/01/2013
County = 40 Oper No. = 777777


Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
06/10/2013 14:55

Test	AC	Time
01 Diagnostic	OK	14:56
02 Room Air	RFI*	14:56
03 Room Air	0.000	14:56

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 61
Expiration Date = 10/01/2013
County = 40 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
06/10/2013 14:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Gas	0.078	14:22
03 Room Air	0.000	14:23
04 Std. Gas	0.079	14:23
05 Room Air	0.000	14:24
06 Std. Gas	0.078	14:24
07 Room Air	0.000	14:25

Lot No = 23411080A1
Cyl No = 61
Exp Date = 10/01/2013
County = 40 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
06/10/2013 14:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:28
02 Std. Gas	0.080	14:29
03 Room Air	0.000	14:29
04 Std. Gas	0.080	14:30
05 Room Air	0.000	14:30
06 Std. Gas	0.080	14:30
07 Room Air	0.000	14:31

Lot No = 23411080A1
Cyl No = 61
Exp Date = 10/01/2013
County = 40 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
06/10/2013 14:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:50
02 Std. Gas	0.080	14:50
03 Room Air	0.000	14:50
04 Std. Gas	0.079	14:51
05 Room Air	0.000	14:51
06 Std. Gas	0.080	14:52
07 Room Air	0.000	14:52


Lot No = 23411080A1

Cyl No = 61

Exp Date = 10/01/2013

County = 40

Oper No. = 777777


Operator Signature
DEB KASHUR

Remarks:

Calibration check 0080A

Form 106-I8000