

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004940

Location: BURK

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>13070</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3841</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5191</u> Lot No. <u>ICS#1</u> Rep AC <u>0.100 AC + 0.05% ACETONE</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>31</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.078</u>	Test 4 <u>0.078</u>	Test 7 <u>0.078</u>
Test 2 <u>0.077</u>	Test 5 <u>0.078</u>	Test 8 <u>0.078</u>
Test 3 <u>0.078</u>	Test 6 <u>0.078</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

L. Tank pressure, Level 3 Func D Sub G P: 830 psi Reg 850 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

<u>17 JUN 13</u> Date	<u>Charles E. Eds</u> Field Inspector's Signature
<u>18 June 13</u> Date	<u>D. Kashner</u> Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
06/17/2013 10:50

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 31
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 720
Intercept -746294

IR Calibration Date: 06/22/2010
 3um 9um

0th Coef(*100):	-10417	-10800
1st Coef(*100):	259947	132173
2nd Coef(*100):	2872	1619
H2O adj(mg/l*10k):	604	506

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
06/17/2013 11:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Sol.	0.016	11:01
03 Room Air	0.000	11:02
04 Std. Sol.	0.016	11:02
05 Room Air	0.000	11:03
06 Std. Sol.	0.017	11:04
07 Room Air	0.000	11:04

08 Sim Temp = 34.0°C

Simul Ser No = DR3803
Std Sol No = 13070
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LOW AC*
 0.020 AC (SIM. SOL.)

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
06/17/2013 11:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Sol.	0.302	11:12
03 Room Air	0.000	11:13
04 Std. Sol.	0.303	11:13
05 Room Air	0.000	11:14
06 Std. Sol.	0.305	11:15
07 Room Air	0.000	11:15


08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 13020

County = 07

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC (SIM. SOL.)

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
06/17/2013 10:55

Test	AC	Time
01 Room Air	0.000	10:56
02 *Subject Test	INT*	10:57
03 Room Air	0.000	10:57

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJH0987
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 31
Expiration Date = 10/01/2013
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
06/17/2013 10:52

Test	AC	Time
01 Diagnostic	OK	10:54
02 Room Air	RFI*	10:54
03 Room Air	0.000	10:54

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 31
Expiration Date = 10/01/2013
County = 07 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
06/17/2013 11:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Gas	0.078	11:18
03 Room Air	0.000	11:19
04 Std. Gas	0.077	11:19
05 Room Air	0.000	11:20
06 Std. Gas	0.078	11:20
07 Room Air	0.000	11:20

Lot No = 23411080A1
Cyl No = 31
Exp Date = 10/01/2013
County = 07 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)
Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
06/17/2013 11:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.078	11:22
03 Room Air	0.000	11:22
04 Std. Gas	0.078	11:23
05 Room Air	0.000	11:23
06 Std. Gas	0.078	11:23
07 Room Air	0.000	11:24

Lot No = 23411080A1
Cyl No = 31
Exp Date = 10/01/2013
County = 07 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
 0.080 AC (E.G.S.)

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
06/17/2013 11:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:25
02 Std. Gas	0.078	11:25
03 Room Air	0.000	11:25
04 Std. Gas	0.078	11:26
05 Room Air	0.000	11:26
06 Std. Gas	0.078	11:27
07 Room Air	0.000	11:27

Lot No = 23411080A1
Cyl No = 31
Exp Date = 10/01/2013
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000