

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004950

Location: WILL

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>13070</u> Rep AC <u>0-020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3841</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5191</u> Lot No. <u>ICS#1</u> Rep AC <u>0.100AC + 0.05% ACETONE</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>50</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.078</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.078</u>	Test 9 <u>0.078</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 187 psi Reg 200 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION
17 JUN 13 Date Charles E. [Signature] Field Inspector's Signature
18 June 13 Date Deb Kasher Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
06/17/2013 15:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 50
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 774
Intercept -908904

IR Calibration Date: 04/06/2011
 3um 9um

0th Coef(*100): -2573 -7967
1st Coef(*100): 257525 132746
2nd Coef(*100): 5623 2169
H2O adj(mg/l*10k): 540 597

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
06/17/2013 15:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:45
02 Std. Sol.	0.019	15:46
03 Room Air	0.000	15:46
04 Std. Sol.	0.019	15:47
05 Room Air	0.000	15:47
06 Std. Sol.	0.020	15:48
07 Room Air	0.000	15:49

08 Sim Temp = 34.0°C

Simul Ser No = DR3803
Std Sol No = 13070
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC (Sim. Sol.)

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
06/17/2013 15:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:51
02 Std. Sol.	0.300	15:52
03 Room Air	0.000	15:53
04 Std. Sol.	0.304	15:53
05 Room Air	0.000	15:54
06 Std. Sol.	0.305	15:55
07 Room Air	0.000	15:55


08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 13020

County = 53

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC (Sim. Sol.)

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
06/17/2013 15:39

Test	AC	Time
01 Room Air	0.000	15:40
02 *Subject Test	INT*	15:41
03 Room Air	0.000	15:41

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJH0987
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 50
Expiration Date = 10/01/2013
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: INTERFERENT CHECK
0.100AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
06/17/2013 15:36

Test	AC	Time
01 Diagnostic	OK	15:37
02 Room Air	RFI*	15:37
03 Room Air	0.000	15:38

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 50
Expiration Date = 10/01/2013
County = 53 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
06/17/2013 15:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:58
02 Std. Gas	0.078	15:58
03 Room Air	0.000	15:59
04 Std. Gas	0.078	15:59
05 Room Air	0.000	16:00
06 Std. Gas	0.079	16:00
07 Room Air	0.000	16:00

Lot No = 23411080A1
Cyl No = 50
Exp Date = 10/01/2013
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
06/17/2013 16:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:01
02 Std. Gas	0.079	16:02
03 Room Air	0.000	16:02
04 Std. Gas	0.079	16:02
05 Room Air	0.000	16:03
06 Std. Gas	0.078	16:03
07 Room Air	0.000	16:04

Lot No = 23411080A1
Cyl No = 50
Exp Date = 10/01/2013
County = 53 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC (E.G.S.)


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
06/17/2013 16:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:05
02 Std. Gas	0.079	16:05
03 Room Air	0.000	16:05
04 Std. Gas	0.079	16:06
05 Room Air	0.000	16:06
06 Std. Gas	0.078	16:07
07 Room Air	0.000	16:07

Lot No = 23411080A1
Cyl No = 50
Exp Date = 10/01/2013
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000