

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004949

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3803</u> Lot No. <u>13070</u> Rep AC <u>0.020</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3841</u> Lot No. <u>13020</u> Rep AC <u>0.300 AC</u>                            | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3846</u> Lot No. <u>—</u> Rep AC <u>0.100 AC</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>23411080A1</u>	Cyl No. <u>68</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 420 psi Reg 440 psi

M. Remarks/Maintenance record: Annual inspection, replaced external battery and o-ring on breath hose!

21 June 13  
Date

Deborah Kashner  
Field Inspector's Signature

21 JUN 13  
Date

Charles E. Eden  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004949  
Location = ~~BELE TOXL~~            8164.13.00 06/09  
06/21/2013                            DK                            13:55

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    23411080A1  
Standard Cyl #?                    68  
Standard Expiration?                10/01/2013  
Oper No?                            777777

Flow Cal. Date:                    09/24/2010  
Slope                                784  
Intercept                            -862571

IR Calibration Date:                09/24/2010  
    3um                                9um

-----  
0th Coef(\*100):                    -23764                                -22121  
1st Coef(\*100):                    267615                                132337  
2nd Coef(\*100):                    1069                                    992  
H2O adj(mg/l\*10k):                531                                    480

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = ~~BELC~~ *TOXL*      8164.13.00 06/09  
06/21/2013      *DK*      13:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:49
02 Std. Sol.	0.018	13:50
03 Room Air	0.000	13:51
04 Std. Sol.	0.019	13:51
05 Room Air	0.000	13:52
06 Std. Sol.	0.019	13:53
07 Room Air	0.000	13:53

08 Sim Temp = 34.0°C

Simul Ser No = DR3803  
Std Sol No = 13070  
County = 08      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Low AC      0.020AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-004949  
Location = ~~BELC~~ 8164.13.00 06/09  
06/21/2013 <sup>TOXL</sup> 13:43  
                  <sub>OK</sub>

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:44
02 Std. Sol.	0.299	13:45
03 Room Air	0.000	13:45
04 Std. Sol.	0.300	13:46
05 Room Air	0.000	13:47
06 Std. Sol.	0.299	13:47
07 Room Air	0.000	13:48

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 13020

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-004949  
Location = ~~BELG~~ *TOXL* 8164.13.00 06/09  
06/21/2013 *DK* 13:55

Test	AC	Time
01 Room Air	0.000	13:56
02 *Subject Test	INT*	13:57
03 Room Air	0.000	13:57

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female Weight = 150  
Test = DUI Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 68  
Expiration Date = 10/01/2013  
County = 08 Oper No. = 777777

*Deb Kasher*  
Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004949  
Location = ~~BELC~~ **TOXL**        8164.13.00 06/09  
06/21/2013                        **DK**                        14:03

Test	AC	Time
01 Diagnostic	OK	14:03
02 Room Air	RFI*	14:04
03 Room Air	0.000	14:04

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                        Weight = 150  
Test = DUI                                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 68  
Expiration Date = 10/01/2013  
County = 08                                        Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = ~~BELC~~ <sup>TOXL</sup>      8164.13.00 06/09  
06/21/2013      <sup>DK</sup>      14:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Gas	0.079	14:05
03 Room Air	0.000	14:06
04 Std. Gas	0.079	14:06
05 Room Air	0.000	14:07
06 Std. Gas	0.079	14:07
07 Room Air	0.000	14:08

Lot No = 23411080A1

Cyl No = 68

Exp Date = 10/01/2013

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = ~~BELG~~ *TOXL*      8164.13.00 06/09  
06/21/2013      *DK*      14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Gas	0.079	14:09
03 Room Air	0.000	14:09
04 Std. Gas	0.079	14:10
05 Room Air	0.000	14:10
06 Std. Gas	0.079	14:11
07 Room Air	0.000	14:11

Lot No = 23411080A1  
Cyl No = 68  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080Ac*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = ~~BELT~~ <sup>TOXL</sup>      8164.13.00 06/09  
06/21/2013      <sup>DK</sup>      14:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Gas	0.079	14:12
03 Room Air	0.000	14:13
04 Std. Gas	0.080	14:13
05 Room Air	0.000	14:14
06 Std. Gas	0.079	14:14
07 Room Air	0.000	14:15

Lot No = 23411080A1  
Cyl No = 68  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080AC