

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004940

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5144</u> Lot No. <u> </u> Rep AC <u>0.021</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5114</u> Lot No. <u> </u> Rep AC <u>0.298</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5186</u> Lot No. <u> </u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 709352 Cyl No. 44 Exp. Date 12/23/2011

Test 1 0.079 Test 4 0.081 Test 7 0.081

Test 2 0.080 Test 5 0.079 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 92 psi Reg 100 psi

M. Remarks/Maintenance record: Annual inspection, recalibration

22 June 2010
Date

Deb Kashur
Field Inspector's Signature

22 Jun 2010
Date

Maury Pearson
Reviewed by

80-004940
p. 34 of 42
JFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
06/22/2010 10:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 44
Standard Expiration? 12/23/2011
Oper No? 777777

Flow Cal. Date: 01/18/2010
Slope 720
Intercept -746294

IR Calibration Date: 06/22/2010

	3um	9um
0th Coef(*100):	-10417	-10800
1st Coef(*100):	259947	132173
2nd Coef(*100):	2872	1619
H2O adj(mg/l*10k):	604	506

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print Test

Form 106-I8000

80-004940
P. 35 of 42
FK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
06/22/2010 10:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:17
02 Std. Sol.	0.019	10:19
03 Room Air	0.000	10:19
04 Std. Sol.	0.020	10:20
05 Room Air	0.000	10:21
06 Std. Sol.	0.020	10:21
07 Room Air	0.000	10:22

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.021 AC

Form 106-I8000

80-004940
P. 36 of 42
pkh

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
06/22/2010 10:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:26
02 Std. Sol.	0.302	10:35
03 Room Air	0.000	10:36
04 Std. Sol.	0.302	10:37
05 Room Air	0.000	10:37
06 Std. Sol.	0.303	10:38
07 Room Air	0.000	10:38

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.298 AC

Form 106-I8000

80-004940
P. 37 of 42
JK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
06/22/2010 15:36

Test	AC	Time
01 Room Air	0.000	15:36
02 *Subject Test	INT*	15:37
03 Room Air	0.000	15:37

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 + acetone

Form 106-I8000

80-004940
p. 38 of 42
KH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
06/22/2010 15:38

Test	AC	Time
01 Diagnostic	OK	15:39
02 Room Air	RFI*	15:39
03 Room Air	0.000	15:40

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-004940
P. 39 of 42
SKH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
06/22/2010 09:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:58
02 Std. Gas	0.079	09:58
03 Room Air	0.000	09:59
04 Std. Gas	0.080	09:59
05 Room Air	0.000	10:00
06 Std. Gas	0.080	10:00
07 Room Air	0.000	10:01

Lot No = 709352
Cyl No = 44
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004940
P. 40 of 42
JK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
06/22/2010 10:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:01
02 Std. Gas	0.081	10:02
03 Room Air	0.000	10:02
04 Std. Gas	0.079	10:03
05 Room Air	0.000	10:03
06 Std. Gas	0.080	10:04
07 Room Air	0.000	10:04

Lot No = 709352
Cyl No = 44
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004940
P. 41 of 42
SEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
06/22/2010 10:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Gas	0.081	10:07
03 Room Air	0.000	10:07
04 Std. Gas	0.080	10:08
05 Room Air	0.000	10:08
06 Std. Gas	0.080	10:09
07 Room Air	0.000	10:09

Lot No = 709352
Cyl No = 44
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004940
p. 42 of 42
KAL