

CMI Calibration Laboratory

Certificate of Calibration

This is to certify the breath alcohol measuring ins after completion of the calibration was found to mand precision. This calibration was performed us <i>International System of Units (SI)</i> traceable through (NIST).	eet or exceed the a	acceptance dards and	e criteria for accuracy materials which are
Calibration Record <u>CR-80-005</u>	952-161865I	<u> </u>	
Estimated Measurement Uncertainty	☑ Original		
Breath Alcohol concentrations of:			
0 ≤ 0.100 g/210L ± <u>0. ○ ○ 2</u> ○ g/210L BAC	☐ Reproduction	Initials	Date
> 0.100 g/210L ± <u>0.</u> <u>0</u> <u>0</u> <u>2</u> <u>0</u> g/210L BAC	☐ Amended		
		Initials	Date
The expanded uncertainties estimated at the time of the calibrate including both Type A and Type B components and are based approximate confiden	on a minimum coverage	using the roo factor of k=2 o	t sum square method corresponding to an
Calibrated by DIROL Date This certificate shall not be reproduced except in full, without written approval of	a second	oB	INTERNATIONAL
			Breath Alcohol Measuring

ISO/IEC 17025:2005 An ASCLD/LAB-International Accredited Laboratory Since February 07, 2012 Certificate # ALI-007-C

Certificate of Calibration

This is to certify the calibration of **Intoxilyzer** serial number 80-05952, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58, No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date June 22, 2013 Signed Del Robrar

Technician



316 East 9th Street Owensboro, KY 42303 USA

Part No. 650517 Rev.A