



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (1-2010)

Serial Number: 80-004947 Instrument Location: OKPD

Reason for Install/Repair:

Install After Lab Repair/Inspection

Other (Specify) Machine Install

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

[Signature]
 Field Inspector Signature

6-23-10
 Date

Deb Kashur
 Reviewed By

25 June 2010
 Date

80-00 4947
 P. 32 of 40
 JAH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.13.00 06/09
06/23/2010 14:29

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 23
Standard Expiration? 09/03/2011
Oper No? 068332

Flow Cal. Date: 01/19/2010
Slope 679
Intercept -681178

IR Calibration Date: 01/19/2010

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -16742 | -27600 |
| 1st Coef(*100): | 254292 | 132484 |
| 2nd Coef(*100): | 3137 | 1165 |
| H2O adj(mg/l*10k): | 754 | 569 |

***** Printer Test End *****



Operator Signature
DAVID M. WALLACE

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.13.00 06/09
06/23/2010 14:29

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:30 |
| 02 Std. Gas | 0.078 | 14:30 |
| 03 Room Air | 0.000 | 14:31 |
| 04 Std. Gas | 0.078 | 14:31 |
| 05 Room Air | 0.000 | 14:32 |
| 06 Std. Gas | 0.079 | 14:32 |
| 07 Room Air | 0.000 | 14:32 |

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 068332



Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004947
P. 39 of 40
JFM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.13.00 06/09
06/23/2010 14:36

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | RFI* | 14:37 |
| 02 Room Air | 0.000 | 14:38 |

*Invalid Test
Inhibited - RFI

Sub Name = DOE, JOHN NA
Sub DOB = 06/06/1960
Sub Sex = Male Weight = 245
Test = DUI Cit = 123456
Dr. Lic. = ND/999999999
Lot No = 659358
Cyl No = 23
Expiration Date = 09/03/2011
County = 45 Oper No. = 068332



Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004947
P. 10 of 40
JDM