



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	<i>David M Wallace</i>		
Location:	<i>Dickinson JEC</i>		
Intoxilyzer® Serial Number:	<i>80 004947</i>		
Gas Lot Number:	<i>659358</i>	Gas Cylinder Number:	<i>23</i>
		Expiration Date:	<i>9-3-11</i>
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)			
Test 1: <i>0.028</i> %			
Test 2: <i>0.029</i> %			
Test 3: <i>0.029</i> %			
(ATTACH TEST RECORD)			

DM Wallace
Chemical Test Operator Signature

6-23-10
Date

Deb Kashur
Reviewed By

25 June 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.13.00 06/09
06/23/2010 14:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:41
02 Std. Gas	0.078	14:41
03 Room Air	0.000	14:41
04 Std. Gas	0.079	14:42
05 Room Air	0.000	14:42
06 Std. Gas	0.079	14:43
07 Room Air	0.000	14:43

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 068332


Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-003 4947
P. 36 of 36
JSM