



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David M Wallace</u>		
Location: <u>Dickinson LEC</u>		
Intoxilyzer® Serial Number: <u>80 - 004947</u>		
Gas Lot Number: <u>65 2358</u>	Gas Cylinder Number: <u>23</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

DM Wallace  
Chemical Test Operator Signature

6-23-10  
Date

Deb Kashur  
Reviewed By

08 July 2010  
Date

80-004947  
P. 43, 44  
FTM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = DKPD      8164.13.00 06/09  
06/23/2010      14:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:51
02 Std. Gas	0.079	14:51
03 Room Air	0.000	14:52
04 Std. Gas	0.079	14:52
05 Room Air	0.000	14:53
06 Std. Gas	0.079	14:53
07 Room Air	0.000	14:54

Lot No = 659358  
Cyl No = 23  
Exp Date = 09/03/2011  
County = 45      Oper No. = 068332



Operator Signature  
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004947  
p. 44 of 44  
KM