

## **CMI** Calibration Laboratory

## **Certificate of Calibration**

This is to certify the breath alcohol measuring instafter completion of the calibration was found to measure and precision. This calibration was performed usinternational System of Units (SI) traceable through (NIST).	eet or exceed the a ing reference stand	acceptance cr dards and ma	iteria for accuracy terials which are
Calibration Record <u>CR-80 -005</u>	950-161865D	)	
Estimated Measurement Uncertainty  Breath Alcohol concentrations of:	□ Original		
	☐ Reproduction		
$0 \le 0.100 \text{ g/}210\text{L} \pm \underline{0}$ , $0 \ge 0$ g/210L BAC	L Reproduction	Initials	Date
> 0.100 g/210L ± 0. 002 4 g/210L BAC	☐ Amended	 Initials	Date
The expanded uncertainties estimated at the time of the calibrati including both Type A and Type B components and are based approximate confiden	on a minimum coverage	using the root sur	n square method
Calibrated by Dir Rok and Date	V 24.20	·	

This certificate shall not be reproduced except in full, without written approval of CMI Calibration Laboratory.



Breath Alcohol Measuring Instrument Calibration ISO/IEC 17025:2005 An ASCLD/LAB-International Accredited Laboratory Since February 07, 2012 Certificate # ALI-007-C

## **Certificate of Calibration**

This is to certify the calibration of **Intoxityzer** ® serial number 80-005950, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58 No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date June 24, 2013 Signed Dol Rohie

316 East 9th Street Owensboro, KY 42303

Part No. 650517 Rev.A