

CMI Calibration Laboratory

Certificate of Calibration

This is to certify the breath alcohol measuring instanter completion of the calibration was found to mand precision. This calibration was performed using international System of Units (SI) traceable through (NIST).	eet or exceed the a	acceptance call dards and ma	riteria for accuracy terials which are
Calibration Record <u>CR-80-0059</u>	58-161865D		
Estimated Measurement Uncertainty Breath Alcohol concentrations of:			
0 ≤ 0.100 g/210L ± <u>0. ∩ ∩ 2 ∩</u> g/210L BAC	☐ Reproduction	Initials	Date
> 0.100 g/210L ± 0. 0 19 g/210L BAC	☐ Amended	Initials	Date
The expanded uncertainties estimated at the time of the calibrati including both Type A and Type B components and are based approximate confiden	on a minimum coverage	using the root sur factor of k=2 corn	m square method esponding to an

This certificate shall not be reproduced except in full, without written approval of CMI Calibration Laboratory.



Breath Alcohol Measuring Instrument Calibration ISO/IEC 17025:2005 An ASCLD/LAB-International Accredited Laboratory Since February 07, 2012 Certificate # ALI-007-C

Calibrated by \(\int \mathbb{\mathbb{\text{\text{\text{Calibrated}}}} \)

Certificate of Calibration

This is to certify the calibration of **Interityzer** ® serial number 60-65958, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58 No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date June 24, 2013 Signed Del Robins

Technician

INC.

316 East 9th Street Owensboro, KY 42303

Part No. 650517 Rev.A