



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	David J. Zeis				
Location:	Cavaher County Sher. P's Office - Langdon				
Intoxilyzer® Serial Number:	80-004188				
Gas Lot Number:	659358	Gas Cylinder Number:	53	Expiration Date:	09/03/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.075 %					
Test 2: 0.076 %					
Test 3: 0.076 %					
(ATTACH TEST RECORD)					

David J. Zeis
Chemical Test Operator Signature

06/25/10
Date

Deb Kashur
Reviewed By

29 June 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = LANG 8164.13.00 06/09
06/25/2010 11:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Gas	0.075	11:11
03 Room Air	0.000	11:12
04 Std. Gas	0.076	11:12
05 Room Air	0.000	11:13
06 Std. Gas	0.076	11:13
07 Room Air	0.000	11:13

Lot No = 659358
Cyl No = 53
Exp Date = 09/03/2011
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZEIS

Remarks:

Form 106-I8000

80-004188
P. 90 of 90
JAN