



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59281 (03-2014)

Serial Number <i>80-005363</i>	Instrument Location <i>Dickinson PD - LEC</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): *NA*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>6-25-15</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>02 July 15</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005363  
Location = DKPD                8164.13.00 06/09  
06/25/2015                      17:44

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            29614080A2  
Standard Cyl #?            10  
Standard Expiration?       12/05/2016  
Oper No?                    010107

Flow Cal. Date:            08/19/2011  
Slope                       660  
Intercept                  -560671

IR Calibration Date:        04/08/2015  
                                    3um                      9um

-----  
0th Coef(\*100):            -12875                      -23244  
1st Coef(\*100):            269300                      140223  
2nd Coef(\*100):            4133                        1487  
H2O adj(mg/l\*10k):        341                         408

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHRISTOPHER MESSER

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005363  
Location = DKPD      8164.13.00 06/09  
06/25/2015      17:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:46
02 Std. Gas	0.078	17:47
03 Room Air	0.000	17:47
04 Std. Gas	0.078	17:47
05 Room Air	0.000	17:48
06 Std. Gas	0.078	17:48
07 Room Air	0.000	17:49

Lot No = 29614080A2  
Cyl No = 10  
Exp Date = 12/05/2016  
County = 45      Oper No. = 010107

  
Operator Signature  
CHRISTOPHER MESSER

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005363  
Location = DKPD                8164.13.00 06/09  
06/25/2015                        17:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:51
02 Std. Gas	0.078	17:51
03 Room Air	0.000	17:52
04 Std. Gas	0.078	17:52
05 Room Air	RFI*	17:52
06 Room Air	0.000	17:53

\*Invalid Test  
Inhibited - RFI

Lot No = 29614080A2  
Cyl No = 10  
Exp Date = 12/05/2016  
County = 45                      Oper No. = 010107

  
Operator Signature  
CHRISTOPHER MESSER

Remarks: