CMI INC.

316 E. 9th St., Owensboro KY 42303
Phone 866-835-0690
Fax: 270-685-6268

SERVICE WORK ORDER

329502

DATE	6-26-12
RCVD VIA	FEDX

BILL TO:			SHIP TO: ND Crime lab		
ATTN:			ATTN: Deb Kashur		
			2641 E. Main Ave.		
CITY	STATE	ZIP	CITY Bismarck	STATE ND	ZIP 58501
CUSTOMER NO. 585013			PHONE 701-328-6159		
PHONE			FAX 701-328-6485		

MODEL #	SERIAL #		BILL CODE
8000	80-004735		

EXTRA PARTS RCVD
AC, Car, Tray, Battery

DESCRIPTION OF PROBLEM
Battery does not stay charged.

Infrared / Fuel Cell
PRELIMINARY TEST INFORMATION

BATTERY- EXT ____ V; INT 7.8 V

Battery pack needs replaced.

VERIFIED PROBLEM? Yes.

WORK PERFORMED:
Replaced power supply & external battery pack. Replaced O-rings on breath tube and simulated return port.

Miscellaneous
PRELIMINARY TEST INFORMATION

Instrument will not turn on.

WORK PERFORMED: Replaced power supply.

PJT

PARTS USED

QTY	PART NO.	DESCRIPTION
1	120094	PSU 12V, 80W
1	112010	Battery 12V NIMH
1	471201	O-ring Sin
1	402565	O-ring Breath

REPAIR	CAL. ADJ.	CALIBRATON	FINAL
MV	DR	DR	DR
½ HOURS 3	HOURS 1	HOURS 5	HOURS 5
DATE July 09, 2012	DATE July 13, 2012	DATE July 13, 2012	DATE July 13, 2012

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS

FORM# CMI SWO REV (1)