CMI INC.

316 E. 9th St., Owensboro KY 42303
Phone 866-835-0690
Fax: 270-685-6268

SERVICE WORK ORDER

329501

DATE 6-26-12
RCVD VIA FEDX

BILL TO:

ATTN:

CITY STATE ZIP

CUSTOMER NO. 585013

PHONE

SHIP TO: N.D. Crime Lab

ATTN: Deb Kashur

3641 E. Main Ave.

CITY Bismarck STATE ND ZIP 58501

PHONE 701-328-6159

FAX 701-328-6188

MODEL # 5000 SERIAL # SQ-0074612 BILL CODE

EXTRA PARTS RCVD
PDT cable box, battery

DESCRIPTION OF PROBLEM
Instrument would not power up

Infrared / Fuel Cell
PRELIMINARY TEST INFORMATION

BATTERY- EXT 12.70 V; INT 7.80 V

VERIFIED PROBLEM?

Unit Had Blown
Fuse in AC Line & Power supply

WORK PERFORMED:

Replaced Parts Listed
Cleaned Unit

Test # 64457

Miscellaneous
PRELIMINARY TEST INFORMATION

WORK PERFORMED:

PDT

PARTS USED

QTY	PART NO.	DESCRIPTION
[illegible]	[illegible]	[illegible]
1	120094	Pwr supply, 80W
1	021352 FLT	Filter AC14
1	013010	Battery 12V, 7AH
1	216120	Flow Sensor

REPAIR	CAL. ADJ.	CALIBRATON	FINAL
[illegible] KA	DR	DT	DR
½ HOURS	HOURS	HOURS	HOURS
8	1	.5	.5
DATE	DATE	DATE	DATE
Aug 20, 2012	Aug 21, 2012	Aug 24, 2012	Aug 27, 2012

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS