DATE	6-26-12
RCVD VIA	FEDX

CMI INC.

316 E. 9th St., Owensboro KY 42303
Phone 866-835-0690
Fax: 270-685-6268

SERVICE WORK ORDER

329500

BILL TO:			SHIP TO: N.D Crime lab		
ATTN:			ATTN: Deb Kashur		
			2641 E. Main Ave.		
CITY	STATE	ZIP	CITY Bismarck	STATE ND	ZIP 58501
CUSTOMER NO.	585013		PHONE 701-328-6159		
PHONE			FAX 701-328-6185		

MODEL #	SERIAL #		BILL CODE
8000	50-004954		1

EXTRA PARTS RCVD
Pit, gas box, bubbling

DESCRIPTION OF PROBLEM
system error 18

Infrared / Fuel Cell
PRELIMINARY TEST INFORMATION

BATTERY- EXT 0 V; INT 3.0 V

system error 18
External battery wont hold charge

VERIFIED PROBLEM? Yes

WORK PERFORMED:
Replaced external battery pack.
Replaced parts listed.
Calibrated & final tested.

Miscellaneous
PRELIMINARY TEST INFORMATION

WORK PERFORMED: Cleared system error.

PARTS USED

QTY	PART NO.	DESCRIPTION
1	013010	12v NiMH
1	402565	Oring
1	471201	Oring

REPAIR	CAL. ADJ.	CALIBRATON	FINAL
MW	DR	DR	DR
½ HOURS	HOURS	HOURS	HOURS
5	1	.5	.5
DATE	DATE	DATE	DATE
Aug 24 2012	Aug 27, 2012	Aug 28, 2012	Aug 28, 2012

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS

FORM# CMI SWO REV (1)