

CHAIN OF CUSTODY  
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-006503

Agency being assigned instrument: Crime Lab

Date 28 June 16

Transferred from:

Name (print) \_\_\_\_\_

Signature \_\_\_\_\_

Agency CMI

Transferred to:

Name (print) Deb Shanaver

Signature Deb Shanaver

Agency Crime Lab

Reason for transfer/remarks:

Optical bench was flooded, per Larry Holm, CMI,  
so it is not covered under warranty.  
A bad Guth simulator flooded this intox.

Or Shipped Via:

USPS \_\_\_\_\_ UPS \_\_\_\_\_ FedEx \_\_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_\_ / \_\_\_\_\_ / \_\_\_\_\_ Air Bill No. \_\_\_\_\_