



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Matthew Anderson</u>		
Location: <u>Richmond County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004206</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>52</u>	Expiration Date: <u>09-03-2011</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.080</u> %		
Test 2: <u>0.079</u> %		
Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

Matthew Anderson  
Chemical Test Operator Signature

6/30/10  
Date

Deb Kashner  
Reviewed By

06 July 2010  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004206  
Location = WAHP      8164.13.00 06/09  
06/30/2010      23:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:46
02 Std. Gas	0.080	23:47
03 Room Air	0.000	23:47
04 Std. Gas	0.079	23:48
05 Room Air	0.000	23:48
06 Std. Gas	0.079	23:48
07 Room Air	0.000	23:49

Lot No = 659358  
Cyl No = 52  
Exp Date = 09/03/2011  
County = 39      Oper No. = 020401



Operator Signature  
MATTHEW W. ANDERSON

Remarks:

*Good test 1st ACA*

Form 106-I8000

*80-004206  
P. 99 of 99  
JMN*