



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59281 (1-2010)

Serial Number: 80-004949 Instrument Location: GFAF

Reason for Install/Repair:

Install After Lab Repair/Inspection

Other (Specify) New instrument

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: N/A

Will Pett  
 Field Inspector Signature

30 June 10 DK  
~~16 Jan 10~~  
 Date

Completed install  
 06/30/2010  
 w/RFI

\_\_\_\_\_  
 Reviewed By

\_\_\_\_\_  
 Date




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = GFAF      8164.13.00 06/09  
06/16/2010      11:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Gas	0.077	11:29
03 Room Air	0.000	11:30
04 Std. Gas	0.077	11:30
05 Room Air	0.000	11:30
06 Std. Gas	0.077	11:31
07 Room Air	0.000	11:31

Lot No = 659358  
Cyl No = 39  
Exp Date = 09/03/2011  
County = 18      Oper No. = 120011

  
\_\_\_\_\_  
Operator Signature  
WILLIAM PETTY

Remarks:

2nd ACA

Form 106-I8000

80-004949  
P. 47-48  
KA

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = GFAF      8164.13.00 06/09  
06/30/2010      12:33

Test	AC	Time
01 Diagnostic	OK	12:37
02 Room Air	RFI*	12:37
03 Room Air	RFI*	12:37

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI, NA NA  
Sub DOB = 12/31/1941  
Sub Sex = Unknown      Weight = NA  
Test = DUI      Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 659358  
Cyl No = 39  
Expiration Date = 09/03/2011  
County = 18      Oper No. = 120011

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
WILLIAM PETTY

Remarks:

Form 106-18000

80-004949  
P. 4P-48  
FM