



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Officer Peter Nielsen

Location: WFPD

Intoxilyzer Serial No.: 80-003059

Lot Number: 659357

Cylinder Number: 10

Expiration Date: ~~9-3-11~~ 12/23/2011 ^{DK}


Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 080 %

Test 2: 0. 079 %

Test 3: 0. 080 %

(ATTACH TEST RECORD)


 Chemical Test Operator Signature
 Deb Kashur

7-2-10
 Date Cylinder Changed
 07 July 2010

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-003059
 P. 93 of 94 *SM*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = WFPD 8164.13.00 06/09
07/02/2010 08:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:59
02 Std. Gas	0.080	08:59
03 Room Air	0.000	09:00
04 Std. Gas	0.079	09:00
05 Room Air	0.000	09:01
06 Std. Gas	0.080	09:01
07 Room Air	0.000	09:02

Lot No = 659357
Cyl No = 10
Exp Date = 12/23/2011
County = 09 Oper No. = 039617



Operator Signature
PETER D. NIELSEN

Remarks:

Form 106-18000

80-003059
p. 94 of 94
KM