

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
07/02/2010 17:24

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 14
Standard Expiration? 12/23/2011
Oper No? 049524

Flow Cal. Date: 01/18/2010
Slope 717
Intercept -728279

IR Calibration Date: 02/10/2010

	3um	9um
0th Coef(*100):	4321	-2310
1st Coef(*100):	254996	129019
2nd Coef(*100):	1560	1203
H2O adj(mg/l*10k):	410	499

***** Printer Test End *****



Operator Signature
TROY A. HISCHER

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
07/02/2010 17:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	17:25
02 Room Air	RFI*	17:25

*Invalid Test
Inhibited - RFI

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 09 Oper No. = 049524



Operator Signature
TROY A. HISCHER

Remarks:

Form 106-I8000

80-004946
P. 47 of 47