



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Officer Peter Nielsen

Location: WFPD

Intoxilyzer Serial No.: 80-003059

Lot Number: 659357

Cylinder Number: 10

Expiration Date: ~~9-3-11~~ 12/23/2011 <sup>DK</sup> Correct date <sup>DK</sup> is 9-3-11 <sup>29 Apr 11</sup>


Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 080 %

Test 2: 0. 079 %

Test 3: 0. 080 %

(ATTACH TEST RECORD)

  
 Chemical Test Operator Signature  
 Deb Kashur

7-2-10  
 Date Cylinder Changed  
 07 July 2010

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-003059  
 P. 10 of 106  
 FM  
 80-003059  
 P. 93 of 94  
 FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = WFPD      8164.13.00 06/09  
07/02/2010      08:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:59
02 Std. Gas	0.080	08:59
03 Room Air	0.000	09:00
04 Std. Gas	0.079	09:00
05 Room Air	0.000	09:01
06 Std. Gas	0.080	09:01
07 Room Air	0.000	09:02

Lot No = 659357      DK 29 April  
Cyl No = 10  
Exp Date = ~~12/23/2011~~ 09/03/11  
County = 09      Oper No. = 039617

  
\_\_\_\_\_  
Operator Signature  
PETER D. NIELSEN

Remarks:

Form 106-I8000

80-003059  
P. 106 of 106  
FM

80-003059  
P. 94 of 94  
FM