



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Richard D. Johnson

Location: McLean S.O. Washburn

Intoxilyzer Serial No.: 80-004196

Lot Number: 659357

Cylinder Number: 47

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.078 %

Test 2: 0.079 %

Test 3: 0.079 %

(ATTACH TEST RECORD)

Richard D. Johnson
 Chemical Test Operator Signature

7-2-2010
 Date Cylinder Changed

Deb Kashui
 Reviewed By

06 July 2010
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004196
 P. 89.490
 JRM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = WASH 8164.13.00 06/09
07/02/2010 10:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:54
02 Std. Gas	0.078	10:55
03 Room Air	0.000	10:55
04 Std. Gas	0.079	10:56
05 Room Air	0.000	10:56
06 Std. Gas	0.079	10:57
07 Room Air	0.000	10:57

Lot No = 659357
Cyl No = 47
Exp Date = 09/03/2011
County = 28 Oper No. = 118707


Operator Signature
RICHARD D. JOHNSON

Remarks:

Form 106-I8000

80-004196
P. 90 of 90
JAN