



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Chris Pulver</u>		
Location: <u>Mountain 50</u>		
Intoxilyzer® Serial Number: <u>80-004943</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>34</u>	Expiration Date: <u>12-23-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>078</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Chris Pulver
Chemical Test Operator Signature

7-6-10
Date

Deb Kashner
Reviewed By

13 July 2010
Date

80-004943
4/7/10 SFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
07/06/2010 15:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:11
02 Std. Gas	0.078	15:11
03 Room Air	0.000	15:12
04 Std. Gas	0.079	15:12
05 Room Air	0.000	15:13
06 Std. Gas	0.079	15:13
07 Room Air	0.000	15:13

Lot No = 709352

Cyl No = 34

Exp Date = 12/23/2011

County = 31

Oper No. = 049912



Operator Signature
CHRISTOPHER PULVER

Remarks:

Form 106-I8000

80-004943
4/8/11 KM