



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Shannon Liebold</u>		
Location: <u>Mercer County Sheriff's Dept.</u>		
Intoxilyzer® Serial Number: <u>80-003060</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>43</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	=
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Shannon Liebold  
Chemical Test Operator Signature

7-7-10  
Date

Deb Kashner  
Reviewed By

21 July 2010  
Date

80-003060  
P. 89 of 90  
km


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = STAT      8164.13.00 06/09  
07/07/2010      13:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:02
02 Std. Gas	0.080	13:02
03 Room Air	0.000	13:03
04 Std. Gas	0.081	13:03
05 Room Air	0.000	13:04
06 Std. Gas	0.080	13:04
07 Room Air	0.000	13:05

Lot No = 659357  
Cyl No = 43  
Exp Date = 09/03/2011  
County = 29      Oper No. = 020904



Operator Signature  
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060  
P. 90 of 90  
JAM